

MASS SPECTRUM SWEEPING IN A
PERMANENT MAGNET MASS SPECTROMETER

Yu.V. Brazhnik, V.F. Shkurdoda

Institute of Applied Physics,
Nat. Acad. of Sci. of Ukraine
(58, *Petropavlovskaya Str., Sumy 244030, Ukraine*)

S u m m a r y

It is shown that a multicollector mass spectrometer with permanent magnet can be used to obtain the mass spectrum in a wide range of masses by slightly varying the ion accelerating voltage and positioning ion collectors in the focal plane in a special way.